



SUBJECT AREAS:  
IMAGING AND SENSING  
FREQUENCY COMBS  
OPTICAL METROLOGY  
OPTICAL SENSORS

## **CORRIGENDUM:** Extremely high-accuracy correction of air refractive index using two-colour optical frequency combs

Guanhao Wu, Mayumi Takahashi, Kaoru Arai, Hajime Inaba & Kaoru Minoshima

This Article contains a typographical error in Equation (4), where a constant coefficient is omitted. Equation (4) should read:

$$\Delta D_1 = -11c \cdot \Delta f_{\text{rep}} / f_{\text{rep}}^2$$

SCIENTIFIC REPORTS:  
3 : 1894  
DOI: 10.1038/srep01894  
(2013)

Published:  
30 May 2013

Updated:  
23 July 2014